

Anesthesiology  
57:543, 1982

### Intra-Arterial Thiopental

*To the Editor:*—A solution of 2.5% thiopental recently was injected into a radial artery line. The total dosage of the drug that the patient received was 150–175 mg. There were no sequelae.

Gangrene of the hand, secondary to an intra-arterial injection of thiopental, is a well-known complication. This usually has occurred with a concentration of 5% or more.<sup>1</sup> In *Introduction to Anesthesia: The Principles of Safe Practice*, it is stated: “. . . to our knowledge, gangrene has not been reported following the use of 2.5 per cent thiopental.”<sup>2</sup>

Communication with several anesthesiologists at other hospitals resulted in differing opinions. Most supported the view that there would be no sequelae. Two of these contacted stated that they had heard of gangrene occurring following an intra-arterial injection of 2.5% thiopental.

Anesthesiology  
57:543, 1982

I have been unable to find, in the literature, a documented case of gangrene resulting after an intra arterial injection of 2.5% thiopental. The persisting question is: Does 2.5% thiopental cause gangrene?

ROSS H. TAFF, M.D.  
Staff Anesthesiologist  
Mount Auburn Hospital  
330 Mount Auburn Street  
Cambridge, Massachusetts 02138

#### REFERENCES

1. Stone HH, Donnelly CC: The accidental intra-arterial injection of thiopental. *ANESTHESIOLOGY* 22:995–1006, 1961
2. Dripps RD, Eckenhoff JE, Vandem LD: *Introduction to Anesthesia: The Principles of Safe Practice*. Philadelphia, WB Saunders, 1977, p182

(Accepted for publication June 6, 1982.)

### Safe Anesthetic Gas Pollution Control

*To the Editor:*—A cross connection of the scavenging circuit into the patient circuit has led repeatedly to near fatalities, both in reported cases and those unreported.<sup>1–3</sup> In such instances, the patient circuit is blocked. This problem was addressed specifically by American National Standard Z-79.11 which requires that the size of the scavenging port be incompatible with both the 15-mm and 22-mm anesthesia circuit end fittings or cuffs (respectively pediatric and adult circuits).\*

Prudence suggests that all equipment not in compli-

ance with the strictest ANS Z-79.11 recommendations, be replaced.

RALPH A. MILLIKEN, M.D., M.B.A.  
Adjunct Associate Professor  
New York Medical College  
234 E. 149th St.  
Bronx, New York 10451

#### REFERENCES

1. Tavakoli M, Habeeb A: Two hazards of gas scavenging. *Anesth Analg (Cleve)* 57:286–287, 1978
2. Flowerdew RMM: A hazard of scavenger port design. *Can Anaesth Soc J* 28:481–483, 1981
3. Mann ES, Sprague DH: An easily overlooked Malassembly. *ANESTHESIOLOGY* 56:413–414, 1982

(Accepted for publication June 2, 1982.)

\* American National Standard Institute Z-79.11. *Anesthesia Gas Pollution Control*, New York, American National Standard Institute, 1982.

Anesthesiology  
57:543–544, 1982

### Rapid Sequence Induction in Patients with a Full Stomach

*To the Editor:*—Rapid sequence induction of anesthesia using thiopental–succinylcholine is commonly performed in patients with a full stomach. Muravchick,

Burkett, and Gold<sup>1</sup> have shown in humans that succinylcholine-induced abdominal fasciculations can be associated with a significant increase in the intragastric