

Anesthesiology
59:482, 1983

Intravenous Infusions: Making Life Easy

To the Editor:—The use of vasoactive drugs delivered by infusion pumps during anesthesia is becoming more and more frequent. Guidelines for infusion concentrations and rates have been given by Hug and Kaplan and have been adopted widely.¹ However, use of these concentrations makes it difficult to calculate how much drug is being delivered at a given moment.

Most infusion pumps are calibrated in ml/h. If the pump is set to deliver x ml/h of a drug concentration y $\mu\text{g/ml}$, then: amount of drug delivery per minute = $\frac{xy}{60}$ $\mu\text{g/min}$.

From this equation it easily can be seen that the usual "round number" concentration of drug gives no simple relationship between the pump setting and the amount of drug being delivered. It is clear, though, that if the drug concentration is a multiple of 60, there will be a fixed relationship between the pump setting and drug infusion rate.

As an example, the suggested concentration for sodium nitroprusside is 50–100 $\mu\text{g/ml}$, which results in a drug delivery rate of 0.83 or 1.67 \times pump setting, respectively. This makes for unnecessary mental arithmetic, should it be necessary to know the infusion rate. If, instead, the drug concentration is 60 or 120 $\mu\text{g/ml}$, the amount of drug delivered per minute equals the pump setting or is twice the pump setting, respectively.

The best way of dealing with mathematics in clinical medicine is to get it out of the way at the start. In the case of anesthesia drug infusions, this can be accomplished by rational selection of drug concentrations. While the drugs are titrated to effect, it is always advantageous to know how much drug is being delivered. Table 1 lists suggested dilutions, the resulting concentrations, and drug infusion rates. These assume an infusion pump calibrated in ml/h.

It may seem odd at first to prepare drugs in these concentrations. However, much time and headache is saved when it is desired to know the amount of drug delivered.

Anesthesiology
59:482–483, 1983

A Simple Device for Monitoring the Esophageal Electrocardiogram

To the Editor:—The diagnosis of complex dysrhythmias is often problematic, especially when the dysrhythmia occurs in the setting of tachycardia.¹ The use of an esoph-

TABLE 1. Suggested Dilutions, the Resulting Concentrations and Drug Infusion Rates

Drug	Suggested Dilution	Infusion Rate, $\mu\text{g/min}$
Dopamine (Intropin®)	60 mg in 100 ml = 600 $\mu\text{g/ml}$	10 \times pump setting
Dobutamine (Dobutrex®)	60 mg in 100 ml = 600 $\mu\text{g/ml}$	10 \times pump setting
Sodium nitroprusside (Nipride®)	12 mg/100 ml = 120 $\mu\text{g/ml}$	2 \times pump setting
Nitroglycerin (Tridil™, and others)	12 mg/100 ml = 120 $\mu\text{g/ml}$	2 \times pump setting
Phentolamine (Regitine®)	12 mg/100 ml = 120 $\mu\text{g/ml}$	2 \times pump setting
Phenylephrine (Neo-Synephrine®)	12 mg/100 ml = 120 $\mu\text{g/ml}$	2 \times pump setting
Metaraminol (Aramine®)	12 mg/100 ml = 120 $\mu\text{g/ml}$	2 \times pump setting
Norepinephrine (Levarterenol)	1.2 mg/100 ml = 12 $\mu\text{g/ml}$	0.2 \times pump setting
Epinephrine (Adrenaline®)	0.6 mg/100 ml = 6 $\mu\text{g/ml}$	0.1 \times pump setting
Isoproterenol (Isuprel®)	0.6 mg/100 ml = 6 $\mu\text{g/ml}$	0.1 \times pump setting

THOMAS D. WEBB, M.D.
Assistant Professor
Department of Anesthesiology
UCLA School of Medicine
Los Angeles, California 90024

REFERENCE

- Hug CC, Kaplan JA: Pharmacology—Cardiac Drugs, Cardiac Anesthesia. Edited by Kaplan JA. New York, Grune and Stratton, 1979, pp 45–50

(Accepted for publication April 7, 1983.)

ageal electrocardiogram often can be a useful technique in this clinical setting to help augment the "P" wave on the electrocardiogram.^{2,3} Kates and colleagues recently