

Effect of Nitrous Oxide on the Oxyhemoglobin Dissociation Curve and P_{O_2} Measurements

J. Reddy Kambam, M.D.,* Duncan A. Holaday, M.D.†

Blood from 20 ASA physical status I patients collected before and after induction of anesthesia was used *in vitro* to reexamine the effects of nitrous oxide on the oxyhemoglobin dissociation curve and the response of the oxygen electrode. The preinduction P_{50} values with and without 70% N_2O were 28.4 ± 0.1 mmHg and 26.8 ± 0.1 mmHg, respectively. The postinduction P_{50} values with and without N_2O were unchanged, namely 28.4 ± 0.1 mmHg and 26.8 ± 0.1 mmHg, respectively. Our data confirm the observation that N_2O causes a small elevation of P_{50} ($P < 0.001$) and that this effect is both rapidly inducible and reversible. Our results also indicate that N_2O has no effect on polarographic P_{O_2} measurements. (Key words: Anesthetics, gases; nitrous oxide. Blood, hemoglobin; oxyhemoglobin dissociation; oxygen saturation. Measurement techniques: electrode; oxygen. Oxygen: oxyhemoglobin dissociation (P_{50}); oxygen saturation; tension.)

N_2O IS THE MOST commonly used inhalational anesthetic today. Data on the effects of N_2O on the oxyhemoglobin dissociation have been conflicting. In 1970, Smith *et al.*¹ studied the effects of N_2O on P_{50} and showed that there is a small increase in P_{50} (shift to the right). In 1984, Fournier and Major² reported a 9 mmHg decrease in P_{50} (shift to the left) when blood samples were treated with N_2O . Fournier and Major also indicated that in the presence of N_2O , P_{O_2} values were falsely increased. If N_2O truly interferes with P_{O_2} measurements and decreases P_{50} to 18 mmHg, as reported by Fournier and Major, it could have impact on critically ill patients in whom N_2O is used as an anesthetic adjunct. In addition, this could complicate studies where oxygen consumption, P_{O_2} , and oxygen content of blood are measured in the presence of N_2O . Accordingly, we reexamined the effects of nitrous oxide on the oxyhemoglobin dissociation curve and on polarographic P_{O_2} measurements.

Methods

With the approval of the Committee for the Protection of Human Subjects, we included in our study 20 ASA physical status I patients who were to receive balanced anesthesia with iv fentanyl, 70% N_2O , and 30% O_2 . Ve-

nous blood samples were drawn from each of these patients before anesthesia was induced and 1 h after anesthesia had been maintained with 70% N_2O and 30% O_2 . Each of the preinduction blood samples was equilibrated in an IL 237 Tonometer[®] with: 1) gas mixtures containing 3.5% O_2 , 4% O_2 , and 4.5% O_2 , with 5.6% CO_2 and balance N_2 ; and 2) similar gas mixtures in addition to 70% nitrous oxide for 15 min at 37° C. The samples collected during anesthesia were equilibrated with the same gas mixtures but in reverse order, *i.e.*, those containing N_2O first. These certified gas mixtures were supplied by MG Industries, scientific gases (Valley Forge, PA) and were tested by us using a mass spectrometer (SARA, Chemtron, Medical Division, St. Louis, MO). At the end of each equilibration, total hemoglobin (Hb), oxyhemoglobin saturations, and carboxyhemoglobin (COHb) were measured in a Radiometer OSM2 Hemoximeter[®] (determination of COHb levels were done in Radiometer, OSM2 Hemoximeter[®] by using Radiometer 904212 deoxygenation capillary kit), and blood gases were measured in a Corning[®] 168 pH/blood gas analyzer. The Radiometer OSM2 Hemoximeter[®] and pH, P_{CO_2} , and P_{O_2} electrodes of Corning[®] 168 pH/blood gas analyzer were calibrated before and after each set of measurements. The measured P_{O_2} data were corrected to pH 7.40 using the Severinghaus formula.³ Because the carbon dioxide tension in the gas mixtures was uniformly 40 mmHg and the blood gas measurements were done at 37° C, no P_{CO_2} or temperature corrections were needed. A three-point saturation curve was plotted in the linear portion of the oxyhemoglobin dissociation curve and P_{50} was obtained for each sample.

Before we proceeded with a P_{50} measurement of each blood sample, we equilibrated 2 ml of the blood with 100% N_2O as well as with a gas mixture containing 50% N_2O and balance N_2 for 15 min at 37° C. At the end of each equilibration, blood gases and O_2 saturation were measured.

Statistical analysis was done for all the measured P_{50} data using students paired *t* test.

Results

Results of this study are shown in tables 1 and 2. P_{50} was 26.8 mmHg without N_2O and 28.4 mmHg with N_2O

* Assistant Professor of Anesthesiology.

† Professor of Anesthesiology.

Received from the department of Anesthesiology, Vanderbilt University Medical Center, Nashville, Tennessee 37232. Accepted for publication August 27, 1986. Presented at ASA Annual Meeting, Atlanta, Georgia, October 1983.

Address reprint requests to Dr. Kambam.

TABLE 1. Preinduction Values of Effect of N₂O on P₅₀

Samples (n = 20)	Without N ₂ O		With N ₂ O		P
	Mean	SEM	Mean	SEM	
3.5% O ₂					
P _{O₂} (mmHg)	26.5	0.03	26.6	0.04	>0.8
O ₂ sat %	49.0	0.05	45.0	0.08	<0.001
4.0% O ₂					
P _{O₂} (mmHg)	30.2	0.04	30.1	0.03	>0.8
O ₂ sat %	58.0	0.07	53.0	0.09	<0.001
4.5% O ₂					
P _{O₂} (mmHg)	34.4	0.05	34.2	0.06	>0.8
O ₂ sat %	66.0	0.08	63.0	0.08	<0.001
P ₅₀ (mmHg)	26.8	0.09	28.4	0.09	<0.001

(*P* < 0.001). The effect of N₂O on P₅₀ was rapidly induced (table 1), and this effect of N₂O on P₅₀ was also rapidly reversible as P₅₀ returned promptly to 26.8 mmHg when blood samples were equilibrated with gas mixtures that contained no N₂O (table 2). The measured P_{O₂} values agreed with the P_{O₂} present in the gas mixtures. No oxygen was detected in the blood samples equilibrated with known gas mixtures containing N₂O but no O₂. This ruled out the possibility of erroneous P_{O₂} values measured in our blood gas machine in the presence of N₂O.

All patients had an Hb between 12 and 15 g/dl and COHb of less than 1%.

TABLE 2. Postinduction Values of Effect of N₂O on P₅₀

Samples (n = 20)	With N ₂ O		Without N ₂ O		P
	Mean	SEM	Mean	SEM	
3.5% O ₂					
P _{O₂} (mmHg)	26.5	0.05	26.4	0.04	>0.8
O ₂ sat %	45.0	0.07	49.0	0.04	<0.001
4.0% O ₂					
P _{O₂} (mmHg)	30.2	0.04	30.1	0.05	>0.8
O ₂ sat %	53.0	0.08	58.0	0.06	<0.001
4.5% O ₂					
P _{O₂} (mmHg)	34.1	0.06	34.3	0.04	>0.8
O ₂ sat %	62.0	0.05	66.0	0.07	<0.001
P ₅₀ (mmHg)	28.4	0.09	26.8	0.08	<0.001

Discussion

Our P₅₀ values are in agreement with those published by Smith *et al.*¹ but disagree with those of Fournier and Major.² Although a statistically significant change in P₅₀ is seen during N₂O anesthesia, it is not clinically significant. Possible mechanisms by which N₂O can interfere with the measurement of P_{O₂} include the presence of silver deposition on the polarographic platinum electrode.⁴ Silver deposition on the cathode increases with age and use of the electrode.‡ Unlike halothane, N₂O cannot be reduced on pure platinum or gold surfaces.⁵ N₂O also interferes with P_{O₂} measurements in certain polarographic oxygen analyzers when the battery output is low or uneven or an incorrect battery is used.§

In conclusion, our data indicate that there is a small but statistically significant shift of P₅₀ to the right in the presence of N₂O, and the effect of N₂O on P₅₀ is both rapidly inducible and reversible. There was no demonstrable interference by N₂O with our polarographic measurements of P_{O₂}.

‡ Severinghaus JW, Bradley AF: Blood gas electrodes or what the instructions didn't say. San Francisco Medical Center, Internal Report, 1969. Copenhagen, Denmark: Radiometer A/S Booklet ST59, 1971.

§ Foregger-Air products, Pamphlet XIV: Oxygen analyzer operating and maintenance.

References

1. Smith TC, Colton ET, Behail MG: Does anesthesia alter hemoglobin dissociation? *ANESTHESIOLOGY* 32:5-11, 1970
2. Fournier L, Major D: The effect of nitrous oxide on the oxyhemoglobin dissociation curve. *Can Anaesth Soc J* 31:173-177, 1984
3. Severinghaus JW: Blood gas calculator. *J Appl Physiol* 21:1105-1116, 1966
4. Albery WJ, Brooks WN, Gibson SP, Hahn CEW: An electrode for P_{N₂O} and P_{O₂} analysis in blood and gas. *J Appl Physiol* 45: 637-643, 1978
5. Severinghaus JW, Weiskopf RB, Nishimura M, Bradley AF: Oxygen electrode errors due to polarographic reduction of halothane. *J Appl Physiol* 31:640-642, 1971