

4. Lederman RP, Lederman E, Work BA Jr, McCann DS: The relationship of maternal anxiety, plasma catecholamines, and plasma cortisol to progress in labor. *Am J Obstet Gynecol* 132:495-500, 1978
5. Melzack R: The myth of painless childbirth (The John J. Bonica Lecture). *Pain* 19:321-337, 1984
6. Shnider SM, Wright RG, Levinson G, Roizen MF, Wallis KL, Rolbin SH, Craft JB: Uterine blood flow and plasma norepinephrine changes during maternal stress in the pregnant ewe. *ANESTHESIOLOGY* 50:524-527, 1979
7. Lederman RP, Lederman E, Work BA Jr, McCann DS: Anxiety and epinephrine in multiparous labor: Relationship to duration of labor and fetal heart rate pattern. *Am J Obstet Gynecol* 153:870-877, 1985
8. Hollmen AI, Jouppila R, Jouppila P, Koivula A, Vierola H: Effect of extradural analgesia using bupivacaine and 2-chloroprocaine on intervillous blood flow during normal labour. *Br J Anaesth* 54:837-842, 1982
9. Huch R, Huch A, Lubbes DW: *Transcutaneous P_O₂*. New York, Thieme-Stratton, 1981, p 139

(Accepted for publication April 2, 1987.)

Anesthesiology
67:154, 1987

More on Management of the Difficult Airway

To the Editor:—I read, with interest, the article by Bedger *et al.*¹ outlining the use of a jet-stylet endotracheal catheter for the management of patients with a difficult airway. We have used a similar technique in our institution. In place of the "jet-stylet," however, we use an 18-French Salem Sump tube with the proximal end of the tube cut off. This allows us to connect our jet ventilator directly to the suction port of the Salem Sump tube. This tube has multiple side holes at the distal end, allowing for ventilation at low gas flows per orifice.

Additionally, during changing endotracheal tubes in the Intensive Care Unit, the nasogastric tube can be passed through the endotracheal tube, the endotracheal tube can be removed, and the patient ventilated prior to or during the endotracheal change process. This technique has been described as a means of changing endotracheal tubes damaged during orthognathic procedures.²

Anesthesiology
67:154-155, 1987

The nasogastric tube is inexpensive, readily available, and easy to use if a jet-stylet is unavailable.

LEWIS A. COVELER, M.D.
*Associate Professor and Chief
Ben Taub General Hospital
Department of Anesthesiology
Baylor College of Medicine
Texas Medical Center
Houston, Texas 77030*

REFERENCES

1. Bedger RC Jr, Chang J-L: A jet-stylet endotracheal catheter for difficult airway management. *ANESTHESIOLOGY* 66:221-223, 1987
2. Coveler LA: Anesthesia for orthognathic surgery, *Current Advances in Oral and Maxillofacial Surgery*. Edited by Irby WB, Shelton DW. St. Louis, C. V. Mosby Company, 1986, pp 38-47

(Accepted for publication April 13, 1987.)

Foreign Body from a Tube of Anesthetic Ointment

To the Editor:—The common practice of lubricating endotracheal tubes with water-soluble ointment may pose an additional risk to our patients. Prior to insertion of an endotracheal tube, pre-lubricated with Xylocaine 5% ointment (Astra Pharmaceutical Products Inc., manufactured by MK Laboratories, Inc., lot #511004, expiration date 11/88), light was noted to be reflected off a foreign body. Closer examination revealed that the small plastic cap liner had become delaminated from its backing and was stuck to the ointment. The plastic seal is clear, and not radioopaque, so its detection to and after insertion of the endotracheal tube would be very difficult. Figure 1 shows the defect as discovered. Examination of our supplies revealed the same defect to exist in many other tubes of the same ointment.




FIG. 1. Cap liner embedded in ointment on the cuff of the endotracheal tube.