

## Design of a Mechanical Stimulator for In Vivo Tissue Engineering of a Diarthrosislike Structure

Jet Human, Eelko May, and Just Herder  
*Delft University of Technology*

Jan Herman Kuiper  
*Keele University*

Jill Urban  
*Oxford University*

James Richardson  
*Keele University*

Sureshan Sivananthan  
*South West London Elective Orthopaedic Centre*

Patrick Warnke

*University of Kiel*

This paper proposes the conceptual design of a mechanical stimulator that uses a tissue engineering strategy to develop a diarthrosislike structure in vivo. The adopted design approach is based on a function analysis. The approach has resulted in the design of a stimulator consisting of four components: cages, a compliant four-bar mechanism, a transmission mechanism, and a fixation component. The implanted stimulator is driven by internal body power, particularly by the longitudinal deformation of a skeletal muscle. The compliant mechanism is designed to impose controlled shear and compressive strain to the growing joint construct in order to initiate cartilage formation. The paper emphasizes the conceptual design and its rationale. Evaluation using finite element analysis was performed, which showed that the design meets the technical demands. Titanium prototypes were fabricated for stiffness and endurance testing.

## Covalent Linking of pH-Sensitive Dye to Fumed Silica

Matthew Nielsen and Megan C. Frost  
*Michigan Technological University*

Measuring the blood gases of patients in the intensive care unit or undergoing major surgical procedures in real time would help healthcare providers diagnose and manage base excess and base deficient disorders. Optical fibers provide a platform upon which an intravascular blood gas sensor may be built and has been by various companies. Unfortunately, thrombosis on the sensor surface interferes with the blood gas measurements and also poses the risk of creating emboli. Nitric oxide inhibits platelet adhesion

and activation, which can reduce thrombus formation on the sensor surface. An optically based pH sensor is fabricated as a first step to show that nitric oxide can be used with blood gas sensors to reduce thrombosis and not interfere with the measurements. pH sensors fabricated using glass microscope slides suffered from leaching of the dye from the sol-gel matrix. The dye readily leached out when the dye was in its appropriate protonated or deprotonated form. To reduce the leaching of the dye, methods of covalently linking the dye to fumed silica have been investigated, and one is presented here.