

Self-Ligation Esthetic Brackets with Low Frictional Resistance

John C. Voudouris^a; Christos Schismenos^b; Kresimir Lackovic^c; Mladen M. Kuffinec^d

ABSTRACT

Objective: To test the frictional resistance forces (FRS) generated between several archwires and (1) interactive self-ligating (ISL) brackets and (2) conventionally ligated (CL) brackets.

Materials and Methods: Frictional forces produced between three different archwire combinations and self-ligating (SL) brackets (ceramic and metal-slot or all-metal) and CL brackets (metal or ceramic) were evaluated in a dry environment. The three ISL brackets tested were In-Ovation-C, In-Ovation-R, and Damon 3. The three CL brackets were Mystique with Neo Clip, Clarity, and Ovation. Each bracket was tested with 0.020" SS, 0.019" × 0.025" SS and 0.018" × 0.018" coated SS.

Results: The ISL brackets generally exhibited the lowest frictional forces irrespective of the bracket material and the wire size, and CL brackets exhibited consistently higher frictional forces. Mystique with Neo Clip produced the lowest frictional resistance of all brackets. The In-Ovation-C brackets demonstrated significantly lower frictional resistance than the SL brackets In-Ovation-R and Damon 3 as well as the CL brackets Clarity and Ovation.

Conclusions: The ISL ceramic brackets produced the lowest frictional resistance of all the self-ligating brackets. The CL ceramic brackets produced the greatest friction. (*Angle Orthod.* 2010; 80:188–194.)

KEY WORDS: Ceramic-injection-molding; Interactive self-ligation; Friction reduction

INTRODUCTION

Interactive self-ligation (ISL) ceramic brackets are relatively new and little evidence-based data have reported thus far. Despite this, the frequency of use of ISL brackets is growing exponentially worldwide (Figures 1 and 2).¹ A major reason for this is the increased patient demand for high esthetics in treatment. Consequently, there is a need for evidence-based research on the properties of esthetic, ceramic ISL and the frictional resistance (FRS) of ISL compared with conventionally ligated (CL) brackets. This is particu-

larly true as lower FRS has recently been associated with significant reductions in overall treatment duration.² In addition, another study demonstrated that passive brackets have no improvement during initial alignment compared with conventional brackets.³

Furthermore, ISL (also referred to as a passive-active system) attempts to satisfy two functional replacement criteria of elastomers.⁴ This includes both locking (passive) and seating (active) the archwire into the base of the slot with low functional friction so as to fully express the prescription.^{5,6} In contrast, passive systems do not fully seat the archwire for early full torque and in-out control as advocated by Roth.⁷ The low FRS found with self-ligation is clinically relevant for early alignment and sliding retraction of canines. Light force application may also contribute to the potential prevention of root resorption and anchorage loss.

The FRS is a critical factor in determining the efficiency of biological tooth movement during sliding mechanics. Many factors have been implicated in influencing friction in orthodontic systems, including wire size,^{4,8–12} bracket slot material,^{11,13,14} and method of ligation.^{4,11,15–20} One ISL bracket study demonstrated lower friction than CL brackets irrespective of the sliding angulations with the archwire.²¹ Other studies have found that increasing the archwire dimension is not

^a Clinical Research Scientist, Department of Orthodontics, New York University, College of Dentistry, New York, NY; Private Practice, Toronto, Ontario, Canada.

^b Postdoctoral student, Department of Orthodontics, New York University, College of Dentistry, New York, NY.

^c Private practice, Toronto, Ontario, Canada.

^d Professor and Director, Department of Orthodontics, New York University, College of Dentistry, New York, NY.

Corresponding author: Dr John Voudouris, College of Dentistry, New York University, Department of Orthodontics, 2300 Yonge St, Suite 707, Toronto, Ontario M4P 1E4 Canada (e-mail: jvoud@pathcom.com)

Accepted: March 2009. Submitted: October 2008.

© 2010 by The EH Angle Education and Research Foundation, Inc.




Figure 1. (A) SEM, 20× magnification of a side view of an inter-active bracket (In-Ovation-C*). A smooth, glasslike texture is observed with the new ceramic-injection-molded (CIM) processed slot. The closed clip is compressed producing a curved shape. This produces a passive-active clip that curves toward the depth of the 0.022" × 0.028 archwire slot to actively seat wires that are ≥0.020" round SS for controlled sliding. Chamfered slot end(s) to reduce couples (binding) during tipping. (B) Light stereomicroscopy of the facial view of the maxillary left canine In-Ovation-C bracket and clip in an open-slot position demonstrates glasslike slot smoothness. The anodized rhodium-coated clip provides an esthetic, pearlescent color blending esthetically with the CIM.




Figure 2. (A) SEM, 25× magnification of side view of the passive SL system Damon 3** of the maxillary left canine is shown with a polycarbonate upper 1/2, and a metal slot-clip assembly in the lower one-half. In this side view, there is no clip curvature to seat archwires toward the base of the slot making it passive. (B) Light stereomicroscopy is used to examine the Frontal-3/4 perspective of the passive metal clip in the open position demonstrating the metal slot-clip assembly is inserted into the polycarbonate material in the open position.

significant;⁹ and some have found that smaller dimension archwires actually produce the highest FRS friction.¹² Several studies have shown stainless steel (SS) brackets have less FRS during sliding than ceramic

Table 1. Groups of Brackets and Ligation Method

Group	Bracket Type	Material Process and Ligation Method
A	In-Ovation-C*	Polycrystalline alumina including slot ceramic-injection molded, interactive self-ligating
B	Damon 3**	Polycarbonate with metal slot and clip assembly, passive self-ligating
C	Mystique*	Milled polycrystalline alumina, with glass slot and Neo Clip* (passive plastic clip)
D	Clarity***	Milled polycrystalline alumina with metal slot liner and elastomers
E	In-Ovation-R*	Metal-injection-molded (MIM) with a milled slot, interactive self-ligating
F	Ovation*	MIM with a milled slot and elastomers

* In-Ovation-C, In-Ovation-R, Ovation, Mystique, Neo Clip, Spectra Wire are registered trademarks of Dentsply/GAC, Bohemia, NY.

** Damon 2, 3 are registered trademarks of Sybron/ORMCO Glendora, California.

*** Clarity, Transbond XT are registered trademarks of 3M/Unitek St. Paul, Minnesota.

and other esthetic brackets.^{11,13,14,22,23} Finally, investigations of ligation methods have shown that ISL brackets reduces friction to minimal levels resulting in better efficiency compared with CL brackets.^{4,16-18}

The purpose of this study was to evaluate the frictional forces generated between several archwires and ISL brackets made of ceramic, polycarbonate-metal, and all-metal brackets and to compare these frictional forces to metal and ceramic-metal brackets with elastomeric ligatures and ceramic with modified ligatures (Table 1). Three SS wire-size combinations with the six brackets were tested in vitro.

MATERIALS AND METHODS

Sample Sizes

FRS was tested by the same operator in a standard laboratory setting using 0.022" × 0.028" maxillary left canine brackets and a Roth prescription. They were combined with 0.020" SS and 0.019" × 0.025" SS archwires (Highland Metals, San Jose, Calif) and 0.018" × 0.018" esthetic SS archwires (Spectra Wire, Dentsply/GAC, Bohemia, NY) coated with polytetrafluoroethylene (PTFE). The brackets tested were:

ISL

Group A⁴—In-Ovation-C (Dentsply/GAC, Bohemia, NY)

Group E⁴—In-Ovation-R (Dentsply/GAC, Bohemia, NY)

Group B⁴—Damon 3 (Sybron/ORMCO, Glendora, Calif)

CL

Group C⁴—Mystique (Dentsply/GAC, Bohemia, NY)

Group D⁴—Clarity (3M/Unitek, St Paul, Minn)

Group F⁴—Ovation, (Dentsply/GAC, Bohemia, NY)

The three different archwires were held in the Clarity and Ovation brackets with elastomeric ligatures (Dentsply/GAC, Bohemia, NY) and tested in a dry environment. The specimen population totaled 180 bracket and archwire specimen combinations (6 groups × 10 sample brackets × 3 archwire sizes). Each bracket was pulled along a section of a 50-mm wire only once to eliminate the influence of wear. The elastomeric modules were tied with a ligature gun (Straight-Shooter, T-P Orthodontics, La Porte, Ind). This method limited possible stretching differences between the elastomeric modules. All elastomers, and all Neo Clips for Mystique, were placed over the bracket-archwire assembly and left for 24 hours before testing to allow for relaxation and adaptation.

Measurement Technique

A testing machine (Chatillon, TCD 200, CSC Force Measurement, Agawam, Mass) with a 1 Newton tension load cell was set on a range of 1 lb and calibrated from 0 to 1000 g in this experiment (Figure 3A). The brackets were bonded (Transbond XT, 3M/Unitek, St Paul, Minn) on plastic bracket pedestals, then wiped with alcohol and air dried. Straight wires were inserted into individual brackets (Figure 3B). The upper and lower part of the wire was then fixed in the C-shaped acrylic rod. The C-shape rod was held from the cross-head of the testing machine (Figure 3C).

Each bracket was placed in a custom-designed testing apparatus that was fixed to the base of the testing machine. The plastic bracket pedestal was connected through roller bearing to allow for linear and rotational displacement of each bracket. Thus, the bracket and the wire specimen were perfectly aligned during each run to prevent a notching or binding effect due to excessive angulation of the bracket relative to the wire.

The rate of movement was 12.7 mm/min, and each test was carried out for 1.0 mm to minimize wear and to simulate canine retraction between monthly appointments. Ten nonrepeated evaluations for each bracket-archwire combination were tested. The load cell registered the force values, as measured by a digital force gauge (Model Chatillon DFA Series, ItinScale Co, Brooklyn, NY) and were transmitted to a computer disk as a DC analog voltage signal. The load cell voltage signals were transmitted to the computer file through a data acquisition board and accompanying software (Nexygen FM, Low Force Application Measurement software Version 4.2, ItinScale Co, Brooklyn, NY). The data were entered into a statistical analysis system (Statistics Toolbox for MATLAB, Mathworks Inc, Natick, Mass).




Figure 3. (A) The frictional resistance testing apparatus. (B) The bracket-wire setup is shown. (C) A magnified illustration of the C-shaped rod holding the wire.

Data Analysis

Mean and standard error were calculated from 10 readings of dynamic friction. A nonrepeated analysis of variance was used to analyze differences among groups, and Tukey's test was used to identify differences among means of different groups for a bracket and wire dimension. Significance was set at $\alpha = .01$.

RESULTS

Comparison of Frictional Forces by Bracket Type

The highest statistically significant friction was shown by the 0.020" SS wire and the Clarity bracket (Group D) (mean 72.55g; Table 2). The Ovation bracket (Group F) produced the second-highest friction (mean 62.55g). The Mystique (Group C), In-Ovation-C (Group A), Damon 3 (Group B), and In-Ovation-R (Group E) brackets produced the lowest mean frictional forces (Figure 4). The differences between Clarity and Mystique and the differences between Ovation and In-Ovation-R were statistically significant at ($P < .01$).

For the 0.019" \times 0.025" SS wire, the Clarity (Group D) and Ovation (Group F) brackets produced the greatest friction with mean values of 185.15g and 154.84g, respectively. The differences between these two were not statistically significant. However, the differences between Clarity and Mystique were statistically significant ($P < .01$). In-Ovation-R brackets demonstrated a mean value for FRS of 98.22g which is statistically significantly lower than Ovation and Clarity brackets ($P < .01$). The lowest frictional forces were produced by Mystique with Neo Clip, In-Ovation-C, and Damon 3 brackets with mean values of 1.59g, 8.33g, and 10.90 g, respectively. In-Ovation-C produced less FRS than Damon 3, but these differences were not statistically significant.

For the 0.018" \times 0.018" SS esthetic coated (Spectra*) wire the Clarity bracket had the highest frictional forces (mean, 177.07g). The Ovation bracket produced the second-highest friction (mean, 121.49g). The lowest frictional forces were shown by Mystique with Neo Clip (mean, 0.72g), In-Ovation-C (mean, 1.94g), Damon 3 (mean, 2.29g), and In-Ovation-R (mean, 18.99g). The difference between the In-Ovation-R metal bracket and the Ovation bracket was statistically significant ($P < .01$).

Comparison by Wire Size and Mode of Ligation

The lowest mean frictional force was noted for the 0.020" SS wire, and the highest mean frictional forces were demonstrated for the 0.019" \times 0.025" SS wire. The 0.018" \times 0.018" SS esthetic coated wire generally showed significantly higher friction (up to 2 \times to 5 \times greater) than the 0.020" SS wires. Overall, ISL brackets showed significantly lower FRS whereas elastomeric ligation produced significantly higher FRS (with the exception of Mystique and NeoClip).

DISCUSSION

The elimination of elastomers and metal ligatures ties has dramatically improved chairside efficiency^{4,24} and the potential for good oral hygiene.^{6,25} Additionally,

Table 2. Means and Standard Deviations (SDs) of Frictional Force in Grams^a

Type of Bracket	Type and Size of Arch Wire Alloy					
	0.020" SS		0.019" × 0.025" SS		0.018" × 0.018" PTFE Coated SS	
	Mean	SD	Mean	SD	Mean	SD
In-Ovation-C	0.44	0.23	8.33	6.93	1.94	1.08
Damon 3	0.53	0.37	10.90	6.15	2.29	1.39
Mystique	0.48	0.35	1.59	1.41	0.72	0.28
Clarity	72.55	13.98	185.15	40.41	177.07	27.62
In-Ovation-R	5.11	1.03	98.22	50.27	18.99	16.14
Ovation	62.55	10.82	154.84	32.72	121.49	20.77

^a PTFE, Polytetrafluoroethylene.

ISL brackets result in a gradual decreasing interactivity. Friction is produced from the time of initial archwire insertion, producing light, continuous forces optimal for biological tooth movement.^{4,5} It is important to note that other factors can affect friction.²⁶

The Effect of Bracket Material, Wire Size, and Ligation Method on Kinetic Friction

The results presented here showed that significant differences were found between ISL esthetic and metal brackets. ISL and SL metal brackets produced similar, lower statistically significant kinetic FRS than CL brackets ($P < .01$). This is contrary to the convention-

ally accepted concept that all ceramic brackets have higher friction than metal brackets.

On preliminary observation, the CL ceramic brackets generally exhibited the highest frictional forces and these results agreed with several other studies.^{13,14,26-28} However, from a clinical perspective, the most significant finding was that an ISL bracket with an all-ceramic slot (In-Ovation-C) demonstrated lower friction than an interactive metal self-ligating (SL) bracket (In-Ovation-R). The similar Cr-Co clips within In-Ovation-C showed more freedom within the bracket body compared with In-Ovation-R. The reduction in the curved shape of the In-Ovation-C clip may have also resulted in lower seating forces, which appears to be related to




Figure 4. Mean frictional resistances with standard deviations are shown for ceramic, polycarbonate-metal, and all-metal SL brackets compared with CL ceramic and metal brackets ligated with elastomers and using three different arch wire sizes.

lower friction. In-Ovation-C consistently demonstrated lower friction than a passive ISL bracket system (Damon 3). Random retesting of bracket-wire combinations confirmed these results.

The aforementioned two findings may be attributable to the new ceramic-injected molding technique that produced a smoother, glass-like slot in In-Ovation-C. Additionally, in clinical settings, a ceramic slot is free of the metal oxide micro-corrosions observed when the metal slots are exposed to saliva.²⁹ The chamfers at the ends of the In-Ovation-C slot may have contributed to reduction in wire binding (Figure 1). In-Ovation-C also showed a smoother vertical slot, possibly allowing easier movement of the interactive clip during the initial light seating of the wire toward the slot base. With respect to direct comparisons of In-Ovation-C vs Damon 3, although the differences were not statistically significant at $P = .01$, In-Ovation-C consistently and for all the tested wires, produced lower FRS.

For control of tooth movement, some frictional resistance must be present. Low friction appears to meet this requirement ideally because it produces earlier rotational corrections and faster initial alignment compared with passive systems.⁴ In the middle stages of treatment, interactive friction is used during canine sliding retraction on 0.018" or 0.020" round SS to prevent distal rotation. In the final stage of treatment, active torque control is observed with the ISL brackets.³⁰ Passive systems lack seating action on the archwire toward the base of the slot.³⁰ Based on our testing and clinical observations, the realistic objective appears to be low interactive friction to seat the wire while producing light, continuous forces for optimal biological tooth movement.⁴ The corollary to this is that low interactive friction has the potential to optimize transeptal fiber pull during various tooth movements. From our empirical observations we suggest that the wire seating maximizes skeletal expansion, including after rapid maxillary expansion, when expanded archwires and headgear are applied.

It was expected that an increase in wire size dimension would produce higher frictional forces, as reported in several other studies.^{8,9,13,21,31,32} The 0.020" SS wire was found to be associated with significantly lower levels of FRS followed by the 0.018" × 0.018" esthetic coated wire that may cause "nudging" dental movements.³² The standard deviations found in this friction investigation are consistent with other friction studies. The ISL esthetic and metal brackets generally produced approximately a 10-fold significantly lower frictional force compared with CL metal and esthetic brackets with elastomers on a 0.020" SS. (Our findings support those of previous studies that found that metal

SL brackets generated lower FRS than did CL metal and CL esthetic brackets.^{21,33,34})

CONCLUSIONS

- The ISL brackets produced the lowest FRS irrespective of bracket material and wire size, whereas the CL brackets exhibited the highest frictional forces, except for Mystique with Neo Clip.
- A ceramic ISL bracket In-Ovation-C showed lower friction than all brackets with metal slots tested.
- The wire size associated with the least friction was 0.020" round SS confirming its indication for ISL sliding biomechanics, followed by 0.018" × 0.018" SS esthetic coated and 0.019" × 0.025" SS wires.

REFERENCES

1. Keim RG, Gottlieb EL, Nelson AH, Vogels DS III. 2008 JCO study of orthodontic diagnosis and treatment procedures. Part 2: breakdowns of selected variables. *J Clin Orthod.* 2008;42:609–710.
2. Pandis N, Polychronopoulou A, Eliades T. Self-ligating vs. conventional brackets in the treatment of mandibular crowding: a prospective clinical trial of treatment duration and dental effects. *Am J Orthod Dentofacial Orthop.* 2007;132:208–215.
3. Miles PG, Weyant RJ, Rustveld L. A clinical trial of Damon 2™ vs conventional twin brackets during initial alignment. *Angle Orthod.* 2006;76:480–485.
4. Voudouris JC. Interactive edgewise mechanisms: form and function comparison with conventional edgewise brackets. *Am J Orthod Dentofacial Orthop.* 1997;11:119–140.
5. Voudouris JC. Flow Mechanics and Biological Tooth Movement. In: Kuffinec MM, ed. *Excellence and Efficiency Interactive Twin Self-Ligation: Orthopedic-Muscle and Gnathological Biomechanics in Clinical Orthodontics.* Toronto, On: ISL and Dentofacial Orthopedic Publications Co; 2007:13–17.
6. Voudouris JC. Seven clinical principles of interactive twin mechanisms. *J Clin Orthod.* 1997;31:55–65.
7. Roth RH. Five-year clinical evaluation of the Andrews' straight-wire appliance. *J Clin Orthod.* 1976;10:836–850.
8. Tidy DC. Frictional forces in fixed appliances. *Am J Orthod Dentofacial Orthop.* 1989;96:249–254.
9. Garner LD, Allai WW, Moore BK. A comparison of frictional forces during simulated canine retraction of a continuous edgewise archwire. *Am J Orthod Dentofacial Orthop.* 1986;90:199–203.
10. Andreasen GF, Quevedo FRS. Evaluation of friction forces in the 0.022 × 0.028 edgewise bracket in vitro. *J Biomech.* 1970;3:151–160.
11. Riley JL, Garrett SG, Moon PC. Frictional forces of ligated plastic and metal edgewise brackets [abstract]. *J Dent Res.* 1979;58:98.
12. Baker KL, Nieberg LG, Weimer AD, Hanna M. Frictional changes in force values caused by saliva substitution. *Am J Orthod Dentofacial Orthop.* 1987;91:316–320.
13. Pratten DH, Popli K, Germane N, Gunsolley JC. FRSS of ceramic and stainless steel orthodontic brackets. *Am J Orthod Dentofacial Orthop.* 1990;98:398–403.
14. Tselepis M, Brockhurst P, West VC. The dynamic FRSS between orthodontic brackets and archwires. *Am J Orthod Dentofacial Orthop.* 1994;106:131–138.

15. Nicolls J. Frictional forces in fixed orthodontic appliances. *Dent Pract.* 1968;18:362–366.
16. Hain M, Dhopalkar A, Rock P. The effect of ligation method on friction in sliding mechanics. *Am J Orthod Dentofacial Orthop.* 2003;123:416–422.
17. Berger JA. The influence of the SPEED bracket's self-ligating design on the force levels in tooth movement a comparative in vitro study. *Am J Orthod Dentofacial Orthop.* 1990;97:219–228.
18. Shivapuja PK, Berger JA. Comparative study of conventional ligation and self ligation bracket systems. *Am J Orthod Dentofacial Orthop.* 1994;106:472–480.
19. Echols PM. Elastic ligatures binding forces and anchorage taxation. *Am J Orthod.* 1975;67:219–220.
20. Baccetti T, Franchi L. Friction produced by types of elastomeric ligatures in treatment mechanics with the preadjusted appliance. *Angle Orthod.* 2006;76:211–216.
21. Thorstenson GA, Kusy RP. Resistance to sliding of self-ligating brackets versus conventional stainless steel twin brackets with second-order angulation in the dry and wet (saliva) states. *Am J Orthod Dentofacial Orthop.* 2001;120:361–370.
22. Cha J-Y, Kim K-S, Hwang C-J. Friction of conventional and silica-insert ceramic brackets in various bracket-wire combinations. *Angle Orthod.* 2007;77:100–107.
23. Kusy R, Whitley J. FRSSs of metal-lined ceramic brackets versus conventional stainless steel brackets and development of 3-D friction maps. *Angle Orthod.* 2001;71:364–374.
24. Maijer R, Smith DC. Time savings with self-ligating brackets. *J Clin Orthod.* 1990;24:29–31.
25. Forsberg CM, Brattstrom V, Malmberg E, Nord CE. Ligation wires and elastomer rings: two methods of ligation, and their association with microbial colonization of *Streptococcus mutans* and *Lactobacilli*. *Eur J Orthod.* 1991;13:416–420.
26. Braun S, Bluestein M, Moore BK, Benson G. Friction in perspective. *Am J Orthod Dentofacial Orthop.* 1999;115:619–627.
27. Keith O, Kusy RP, Whitley JQ. Evaluation of zirconia brackets. *Am J Orthod Dentofacial Orthop.* 1994;106:605–614.
28. Kemp DW. *A Comparative Analysis of Frictional Forces between Self-ligating and Conventional Edgewise Orthodontic Brackets* [thesis]. Toronto, Canada: Department of Orthodontics, University of Toronto; 1992.
29. Matasa GC. Attachment corrosion and its testing. *J Clin Orthod.* 1995;29:16–23.
30. Voudouris JC, Kuftinec MM, Bantleon HP, Muhs S, Perscheck A. Selbsligierende Twin-Brackets [Teil I]—ist weniger mehr? *Inf Orthod Kieferorthop.* 2003;35:13–18.
31. Kapila S, Angolkar PV, Duncanson MG, Nanda RS. Evaluation of friction between edgewise stainless steel brackets and orthodontic wires of four alloys. *Am J Orthod Dentofacial Orthop.* 1990;98:117–126.
32. Kusy RP, Whitley JQ, Mayhew MJ, Buckthal JE. Surface roughness of orthodontic arch wires via laser spectroscopy. *Angle Orthod.* 1988;58:33–45.
33. Pizzoni L, Ravnholt G, Melsen B. Frictional forces related to self-ligating brackets. *Eur J Orthod.* 1998;20:283–291.
34. Thorstenson GA, Kusy RP. Comparison of resistance to sliding between different self-ligating brackets with second order angulation in the dry and saliva states. *Am J Orthod Dentofacial Orthop.* 2002;121:472–482.