



User Engagement With the CamAPS FX Hybrid Closed-Loop App According to Age and User Characteristics

Diabetes Care 2021;44:e148–e150 | <https://doi.org/10.2337/dc20-2762>

Natalie S. Chen,¹
 Charlotte K. Boughton,^{1,2} Sara Hartnell,²
 Julia Fuchs,^{1,3} Janet M. Allen,¹
 Malgorzata E. Willinska,¹
 Ajay Thankamony,³ Carine de Beaufort,⁴
 Fiona M. Campbell,⁵
 Elke Fröhlich-Reiterer,⁶ Sabine E. Hofer,⁷
 Thomas M. Kapellen,⁸
 Birgit Rami-Merhar,⁹ Atrayee Ghatak,¹⁰
 Tabitha L. Randell,¹¹
 Rachel E. J. Besser,^{12,13} Daniela Elleri,¹⁴
 Nicola Trevelyan,¹⁵ Louise Denvir,¹¹
 Nikki Davis,¹⁵ Eleanor Gurnell,²
 Robert Lindsay,¹⁶ Damian Morris,¹⁷
 Eleanor M. Scott,¹⁸ Lia Bally¹⁹
 Hood Thabit,²⁰ Lalantha Leelarathna,²⁰
 Mark L. Evans,^{1,2} Helen R. Murphy,²¹
 Julia K. Mader,²² and Roman Hovorka,¹
 on behalf of AiDAPT, AP@home04,
 CLOuD, DAN05, DAN06, and KidsAP
 consortia

CamAPS FX (CamDiab, Cambridge, U.K.) is a hybrid closed-loop app hosting the Cambridge closed-loop algorithm on an Android smartphone, and it is approved in the European Union for use in children ≥ 1 year and adults (including

during pregnancy) with type 1 diabetes (T1D). The interoperable CamAPS FX app receives glucose data from a compatible continuous glucose monitoring system (Dexcom G6; Dexcom, San Diego, CA), connects to a compatible insulin

pump (Dana Diabecare RS and DANA-i; Sooil, Seoul, South Korea) to direct glucose-responsive insulin delivery every 8–12 min, includes a bolus calculator allowing discrete bolusing via the app, and streams data in real time to cloud-based

¹Wellcome Trust-MRC Institute of Metabolic Science, University of Cambridge, Cambridge, U.K.

²Wolfson Diabetes and Endocrine Clinic, Cambridge University Hospitals NHS Foundation Trust, Cambridge, U.K.

³Department of Paediatrics, University of Cambridge, Cambridge, U.K.

⁴Diabetes Endocrinology Care Clinique Pédiatrique, Clinique Pédiatrique, Centre Hospitalier de Luxembourg, Luxembourg

⁵Department of Paediatric Diabetes, Leeds Children's Hospital, Leeds, U.K.

⁶Department of Pediatrics and Adolescent Medicine, Medical University of Graz, Graz, Austria

⁷Department of Pediatrics I, Medical University of Innsbruck, Innsbruck, Austria

⁸Hospital for Children and Adolescents, University of Leipzig, Leipzig, Germany

⁹Department of Pediatrics and Adolescent Medicine, Medical University of Vienna, Vienna, Austria

¹⁰Alder Hey Children's NHS Foundation Trust, Liverpool, U.K.

¹¹Nottingham Children's Hospital, Nottingham, U.K.

¹²Department of Paediatrics, University of Oxford, Oxford, U.K.

¹³NIHR Oxford Biomedical Research Centre, John Radcliffe Hospital, Oxford, U.K.

¹⁴Royal Hospital for Sick Children, Edinburgh, U.K.

¹⁵Southampton Children's Hospital, Southampton, U.K.

¹⁶Glasgow Royal Infirmary, Glasgow, U.K.

¹⁷Ipswich Hospital, Ipswich, U.K.

¹⁸Department of Population and Clinical Sciences, Leeds Institute of Cardiovascular and Metabolic Medicine, University of Leeds, Leeds, U.K.

¹⁹Department of Diabetes, Endocrinology, Clinical Nutrition and Metabolism, Inselspital, Bern University Hospital, Bern, Switzerland

²⁰Diabetes, Endocrinology and Metabolism Centre, Manchester University NHS Foundation Trust, Manchester, U.K.

²¹Norwich Medical School, University of East Anglia, Norwich, U.K.

²²Division of Endocrinology and Diabetology, Department of Internal Medicine, Medical University of Graz, Graz, Austria

Corresponding author: Charlotte K. Boughton, cb2000@medschl.cam.ac.uk

Received 11 November 2020 and accepted 6 April 2021

© 2021 by the American Diabetes Association. Readers may use this article as long as the work is properly cited, the use is educational and not for profit, and the work is not altered. More information is available at <https://www.diabetesjournals.org/content/license>.

diabetes data repositories (Diasend/Glooko, Gothenburg, Sweden).

It is currently unknown how much time different user cohorts spend interacting with hybrid closed-loop systems. We aimed to investigate usage patterns of the CamAPS FX app across different populations by considering the amount of time a person spends using the app over a 24-h period.

We noted time spent within the CamAPS FX app over an 11-week period from February to May 2020 in 134 individuals from six ongoing clinical studies. Each study involved a different demographic cohort: very young children (1–7 years) in which caregivers use the app (NCT03784027, ClinicalTrials.gov), children and adolescents (6–19 years) (NCT02925299), adolescents (10–17 years) using the closed-loop system from diagnosis of T1D (NCT02871089), adults (≥ 18 years) (NCT04055480), pregnant women (≥ 18 years) (ISRCTN56898625, www.isrctn.org), and older adults (≥ 60 years) (NCT04025762). All studies received regulatory and ethical approval; participants/guardians signed informed consent.

User engagement is the duration of time that users have the application active and in the foreground, as calculated by an analytics cloud platform (Firebase; Google). Time spent using the application was divided into nighttime (0000–0559 h) and daytime (0600–2359 h). For each day of the 11-week observation period, average app usage was recorded per study cohort. Data are

presented as median with interquartile range of these daily recordings.

The mean time spent in the CamAPS FX app across all user cohorts was 36 min/day. This includes all aspects of app engagement (initiating prandial insulin boluses, responding to sensor glucose and other alerts, reviewing data, altering app settings for exercise, troubleshooting connectivity issues, and starting/stopping the glucose sensor). Nocturnal engagement was low, with just 3 min of app usage, on average, during nighttime.

Participants from distinct demographic cohorts differed in the amount of time spent within the app, ranging from 10 to 81 min/day (Fig. 1). Caregivers of very young children and older adults had greatest app engagement (median [interquartile range] 81 [63–96] and 63 [39–83] min/day, respectively). Caregivers of very young children had the greatest nighttime engagement, reflecting the sleep disruption and anxiety around monitoring children's nocturnal glucose reported in this population (1).

Children and adolescents with established T1D spent time engaging with the app similar to that of adult users (16 [13–21] vs. 16 [14–18] min/day, respectively), while adolescents using the closed-loop system from diagnosis spent the least amount of time in the app, at 10 [9–11] min/day. Pregnant women, who target tighter glycemic control, spent double the amount of time in app spent by nonpregnant adults (32 [25–40] vs. 16 [14–18] min/day).

Time in app across pediatric cohorts overall was lower than that reported in a survey of diabetes educators, where the estimated time required for routine diabetes self-management for a child with T1D using multiple daily insulin injections and fingerstick glucose monitoring is 78 min/day (2).

A feasibility trial of long-term closed-loop use in adults reported that users spending “low” amounts of time engaging with the system reported fewer benefits than “high” users who used the system for longer periods (3). Increased time spent with closed-loop in operation is associated with improved glycemic outcomes (4). High users must also consider the burden of diabetes management for optimal long-term benefits and prevention of management-related fatigue and burnout. There is a need to balance optimal glucose control and time spent managing diabetes using hybrid closed-loop to truly reduce burden.

Strengths of the study include evaluating usage patterns across different user characteristics from multicenter, and, for some, multinational studies, thereby increasing generalizability. Whether these engagement patterns are similar to those observed with closed-loop systems, where the control algorithm resides on the insulin pump (5), remains to be determined. Limitations include an inability to link app engagement to glucose outcomes and an unbalanced number of subjects per cohort. The duration of time participants had been using the app prior to the observation period varied between studies and individuals within a study.

We conclude that considerable differences in closed-loop app engagement exist among different user cohorts. Further research is warranted to elucidate drivers for user engagement and identify means to reduce overall diabetes burden for people with T1D and caregivers using hybrid closed-loop systems.




Figure 1—Duration of user engagement during nighttime (0000–0559 h) and daytime (0600–2359 h) with the CamAPS FX app across six closed-loop studies.

Funding. This study was supported by the National Institute for Health Research Cambridge Biomedical Research Centre, JDRF, European Union Horizon 2020 research and innovation program (731560), National Institute for Health Research Efficacy and Mechanism Evaluation grants (14/23/09 and 16/35/01), National Institute of Diabetes and Digestive and Kidney Diseases (UC4DK108520), The Leona M. and Harry B. Helmsley

Charitable Trust, and Wellcome Trust Strategic Award (100574/Z/12/Z).

The views expressed are those of the authors and not necessarily those of the funders.

Duality of Interest. S.H. serves as a member of Sigma-Aldrich (Dexcom) and Medtronic advisory boards; is a consultant for CamDiab and a director of Ask Diabetes Ltd., providing training and research support in health care settings; and reports having received training honoraria from Medtronic and Sanofi. M.E.W. reports receiving license fees from B. Braun, patents related to closed-loop systems, and being a consultant at CamDiab. E.F.-R. reports having received speaker honoraria from Medtronic, Eli Lilly and Company, Novo Nordisk, and Sanofi and serving on advisory boards for Eli Lilly and Company. B.R.-M. received speaker honoraria from Eli Lilly and Company, Medtronic, Novo Nordisk, Roche Diabetes Care, and Sanofi. R.E.J.B. reports having received speaking honoraria from Eli Lilly and Company. R.L. reports having received speaker and advisory board fees from Novo Nordisk, Servier, and Eli Lilly and Company. E.M.S. reports having received speaker honoraria from Abbott Diabetes Care and Eli Lilly and Company. H.T. reports having received research support and speaker honoraria from Dexcom. L.L. reports having received speaker honoraria from Animas, Abbott, Insulet, Medtronic, Novo Nordisk, Roche, and Sanofi; was on advisory panels for Animas, Abbott, Novo Nordisk,

Dexcom, Medtronic, Sanofi, and Roche; and received research support from Novo Nordisk and Dexcom. M.L.E. reports having received speaker honoraria from Eli Lilly and Company, Novo Nordisk, Abbott Diabetes Care, Medtronic, AstraZeneca, and Ypsomed and acting on advisory boards for Medtronic, Zucara Therapeutics, Pila Pharma, and Abbott Diabetes Care. J.K.M. is a member of the advisory boards of Abbott Diabetes Care, BD, Boehringer Ingelheim, Eli Lilly and Company, Medtronic, Prediktor A/S, Roche Diabetes Care, and Sanofi; received speaker honoraria from Abbott Diabetes Care, AstraZeneca, Dexcom, Eli Lilly and Company, Novo Nordisk A/S, Roche Diabetes Care, Servier, and Takeda Pharmaceutical Company; and is a cofounder and shareholder of decide Clinical Software Ltd. R.H. reports having received speaker honoraria from Eli Lilly and Company, Dexcom, and Novo Nordisk; receiving license fees from Medtronic; receiving patents related to closed-loop systems; and being director at CamDiab. No other potential conflicts of interest relevant to this article were reported.

Author Contributions. N.S.C., C.K.B., and R.H. codesigned the evaluations. N.S.C. carried out the data analysis. C.K.B., S.H., J.F., J.M.A., M.E.W., A.T., C.d.B., F.M.C., E.F.-R., S.E.H., T.M.K., B.R.-M., A.G., T.L.R., R.E.J.B., D.E., N.T., L.D., N.D., E.G., R.L., D.M., E.M.S., L.B., H.T., L.L., M.L.E., H.R.M., and J.K.M. provided patient care for participants in the

studies. N.S.C., C.K.B., and R.H. wrote the manuscript. All authors critically reviewed the report. No writing assistance was provided. R.H. is the guarantor of this work and, as such, had full access to all of the data in the study and takes responsibility for the integrity of the data and the accuracy of the data analysis.

References

1. Monaghan MC, Hilliard ME, Cogen FR, Streisand R. Nighttime caregiving behaviors among parents of young children with type 1 diabetes: associations with illness characteristics and parent functioning. *Fam Syst Health* 2009;27:28–38
2. Shubrook JH, Brannan GD, Wapner A, Klein G, Schwartz FL. Time needed for diabetes self-care: nationwide survey of certified diabetes educators. *Diabetes Spectr* 2018;31:267–271
3. Kovatchev B, Cheng P, Anderson SM, et al. Feasibility of long-term closed-loop control: a multicenter 6-month trial of 24/7 automated insulin delivery. *Diabetes Technol Ther* 2017; 19:18–24
4. Lal RA, Basina M, Maahs DM, Hood K, Buckingham B, Wilson DM. One year clinical experience of the first commercial hybrid closed-loop system. *Diabetes Care* 2019;42:2190–2196
5. Brown SA, Kovatchev BP, Raghinaru D, et al.; iDCL Trial Research Group. Six-month randomized, multicenter trial of closed-loop control in type 1 diabetes. *N Engl J Med* 2019;381:1707–1717