

Author Response: In Vivo Oxygen Uptake into the Human Cornea

Fink and Hill¹ question the transport of oxygen to the detection probe during in vivo polarographic assessment of corneal oxygen uptake. Polarographic measurement of corneal oxygen uptake is based on electrochemical reduction of oxygen at the surface of a Clark electrode.²



This reaction demands four electrons for every oxygen molecule consumed and, accordingly, gives rise to a cathodic current. Electrical current resulting from the reduction process in reaction 1 is measured by the polarographic instrument.² Without a supply of oxygen to the electrode surface (i.e., a diffusion flux), reaction 1 cannot consume oxygen, no electrical current can arise, and no instrument signal is possible. We reiterate that the polarographic oxygen sensor (POS) directly measures the electrical current arising from the consumption of oxygen following reaction 1 and not oxygen partial pressure.³ If there is no supply of oxygen to the electrode surface, reaction 1 ceases and no measurement can be made. A detectable instrument signal demands a flux of oxygen directed from the membrane towards the electrode. Our approach faithfully represents this well established physical fact, whereas the currently used analysis of Jauregui and Fatt⁴ demands a zero oxygen flux to the electrode. Consequently, the data-interpretation scheme of Fatt and coworkers for the POS is inconsistent with the factual physical behavior of the Clark electrode. The schematics in Figures 1 and 2 simply illustrate this established fact.³

The sample size of our human subject trial was not large enough to assert a definitive value for corneal oxygen uptake. Nevertheless, our value does fall within the range cited for a sample size of 100.⁵ We stand by our

analysis and interpretation scheme for the POS, and we strongly recommend that future measurements be so interpreted.

*Sbo C. Takatori*¹
Percy Lazon de la Jara^{2,3}
Brien Holden^{2,3}
*Klaus Ehrmann*²
Arthur Ho^{2,3}
Clayton J. Radke^{1,4}

¹Department of Chemical and Biomolecular Engineering, University of California, Berkeley, California; the ²Brien Holden Vision Research Institute, Sydney, Australia; the ³School of Optometry & Vision Science, University of New South Wales, Sydney, Australia; and the ⁴Vision Science Group, University of California, Berkeley, California.

E-mail: radke@berkeley.edu

References

1. Fink BA, Hill RM. In vivo oxygen uptake into the human cornea. *Invest Ophthalmol Vis Sci.* 2012;53:6828.
2. Clark LC, Sachs G. Bioelectrodes for tissue metabolism. *Ann NY Acad Sci.* 1968;148:133-153.
3. Takatori SC, Lazon de la Jara P, Holden B, Ehrmann K, Ho A, Radke CJ. In vivo oxygen uptake into the human cornea. *Invest Ophthalmol Vis Sci.* 2012;53:6331-6337.
4. Jauregui MJ, Fatt I. Estimation of oxygen tension under a contact lens. *Am J Optom Arch Am Acad Optom.* 1971;48:210-218.
5. Quinn TG, Schoessler JP. Human corneal epithelial oxygen demand-population characteristics. *Am J Optom Physiol Opt.* 1984;61:386-388.

Citation: *Invest Ophthalmol Vis Sci.* 2012;53:6829.
 doi:10.1167/iovs.12-10901